



Member of the FM Global Group

FM Approvals
1151 Boston-Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781 762 9375 www.fmglobal.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

MK 5090 Analog Flometer.

IS / I,II,III / ABCDEFG / T5

MK 515 Paddlewheel Flow Sensor

IS / I,II,III / ABCDEFG / T5

MK525- Paddlewheel Flow Sensor.

IS / I,II,III / ABCDEFG / T5 — P52565 / D

3-2517- Paddlewheel Flow Sensor.

IS / I,II,III / ABCDEFG / T5 — P52565 / D

MK525-a Paddlewheel Flow Sensor.

XP / I / 1 / BCD / T5; DIP / II,III / 1 / EFG / T5; S / III / 2 / T5

a = Installation fitting 1, 2 or 3.

3-2517-a Paddlewheel Flow Sensor.

XP / I / 1 / BCD / T5; DIP / II,III / 1 / EFG / T5; S / III / 2 / T5

a = Installation fitting 1, 2 or 3.

Equipment Ratings:

Intrinsically Safe electrical apparatus for use in Class I, II, III, Division 1, Groups A, B, C, D, E, F and G when installed per documents P52565; indoor hazardous (classified) Locations

FM Approved for:

George Fischer Signet Inc,
El Monte CA 91734

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3600	1989
Class 3610	1988
Class 3615	1989
Class 3810 Supp No. 1	1995

Original Project ID: 2K3A8.AX


Approval Granted: October 15, 1985

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
2M1A4.AX	April 24, 1986		
2P1A2.AX	June 6, 1988		
2T4A0.AX	July 11, 1991		
4Y8A7.AX	March 4, 1996		
2T4A0.AX	February 9, 2004		

FM Approvals LLC


Robert L. Martell, Jr.
Assistant Vice President


Date